

TEST INSTRUCTION

For stool sample collection with OmicSnap Tube



YOU RECEIVE

The sample tubes may already be inside the transport tubes.



Here you will find the test instruction video:



OPTIONAL




1 x Request form



1 x Transport tube



1 x OmicSnap Tube (with stabilizing buffer + matrix)



1 x Name label



1 x Transport tube



1 x Stool tube

Purpose:


The OmicSnap Tube is an all-in-one system for the manual extraction of a defined amount of stool by laypersons and enables the stabilisation and homogenisation of the sample for further use in nucleic acid-based procedures in the laboratory.

Safety advice:

Do not drink the contents of the stool sample tube, avoid direct contact with the liquid. After skin contact, immediately wash the affected areas thoroughly. Keep out of reach of children. Always wash/disinfect hands after collecting stool samples. Serious incidents must be reported to Immundiagnostik AG and the national supervisory authorities. This sampling set has been manufactured and placed on the market in accordance with Regulation (EU) 2017/746 (IVDR).







 **Manufacturer:**
Immundiagnostik AG
Stubenwald-Allee 8a
64625 Bensheim, Germany
Tel.: +49 6251 70190-0
Fax: +49 6251 70190-363
info@immundiagnostik.com
www.immundiagnostik.com

 **Distribution by:**
biovis Diagnostik MVZ GmbH
Brüsseler Str. 18
65552 Limburg-Eschhofen
Tel.: +49 6431 21248 0
Fax: +49 6431 21248 66
info@biovis.de


biovis.de



 KV 10021 CP
 n-vitro diagnostics
 Batch number
 Unique Device Identifier

 Storage temperature: 2-30°C

 Usable until


 Do not reuse



TEST INSTRUCTIONS

For stool sample collection - OmicSnap Tube


1. Place all the materials in front of you. Place the stool collector on the toilet seat.



IMPORTANT

The stool sample should not come into contact with the toilet bowl or water.

2. Place the OmicSnap tube upright, to prevent the liquid from escaping. Unscrew the yellow cap counterclockwise and remove the sample lance.




IMPORTANT


The white cap must absolutely remain untouched. Please do not open.

3. Insert the sample lance at four different spots into the stool sample, until the sample lance is filled.


Please do not use the sample lance as a spoon to transfer the stool into the sample tube.



4. Remove excess stool from the top of the sample lance.




5. Return the sample lance back into the OmicSnap tube and turn the yellow cap until resistance engages and it cannot be opened.




6. Then swirl the OmicSnap tube back and forth several times

Please do not shake.



Optionally, use the same stool for the second stool tube.

7. Unscrew the stool collection tube (brown lid) and take stool samples from at least four different spots. Fill the stool collection tube until it reaches at least the marked line. Screw the lid with the attached spoon back onto the tube.



8. Remove the stool collector from the toilet seat. It is water-soluble and can be flushed away.



9. The practice should have already labelled the sample tubes with a barcode for you.

If not, write your **name** and the **date or time of sample collection** on the name label and attach it to the OmicSnap tube.

Optionally, label the stool tube with **name, date of birth** and **date of sample collection**.




10. Please fill in the request form legibly (in digital or printed letters).

Write your name, date of birth, and date and time of collection on the front.

Complete the declaration on the back and sign the request form.



11. First pack the sample tubes in the transport tubes and then place them in the mailing bag with the completed and signed request form.



12. Drop off the shipping bag at a post office as soon as possible, or — after prior consultation — at your medical practice.

Please do not send on Fridays, weekends or public holidays and do not drop in letterboxes.

If you have any further questions, please contact your practice.

